



औरंगाबाद शहर आरोग्य समिती महानगरपालिका औरंगाबाद.

(आरोग्य विभाग)



दुरध्वनी क्र.(०२४०) २३३३५३६-४०
(विस्तार) २३९
कार्यालय दुरध्वनी क्र.०२४० २३३८०२८
फॅक्स क्र. (०२४०) २३३१२१३

पत्ता - औरंगाबाद महानगरपालिका, आरोग्य विभाग, मुख्य
कार्यालय, टाऊन हॉल औरंगाबाद ४३१००१
Website - contact@aurangabadmahapalika.org
Email ID - fwsipamcabd@yahoo.co.in

आरोग्य सेवा

जा.क्र.मनपा/आरोग्य/एनयुएचएम/ 10/२०१९ दि.०६/०६/२०१९

दरपत्रक सुचना

प्रति

.....
.....
.....

विषय :- नविन आरोग्य केंद्रा करीता आवश्यक साहित्य खरेदी करण्यासाठी
दरपत्रक सादर करणे.

उपरोक्त विषयी आपणास कळविण्यात येते कि, औरंगाबाद शहर आरोग्य
समिती महानगरपालिका, औरंगाबाद अंतर्गत बांधण्यात आलेल्या नविण चेतना
नगर आरोग्य केंद्रासाठी खालील प्रमाणे साहित्य खरेदी करावयाचे आहे त्याची
यादी खालील प्रमाणे आहे.

PROPOSED NEW UPHC PURCHASE				
ITEM	Qty Proposed	Rate Per	Estimated Rates	
Gyanacology Dept				
I.U.C.D. Set	2	1		
P.V. Tray	2	1		
Suturing set	1	1		
General				
Examination table	1	1		
Hammer	2	1		
Tough Depressor	2	1		
Thermometer Dial	2	1		
Steel dressing jar	2	1		
Steel kidney tray 8"	2	1		
Steel bowl 10"	2	1		
Bed side screen	2	1		

	Stethoscope Pediatric	2	1
	Torch	2	1
	Thermometer Clinical	2	1
	Dressing drum 6"x 6"	2	1
	Dressing drum 9"x 9"	2	1
	Steel bed side locker	2	1
	Steel instrument trolley	1	1
	Steel I V stand	2	1
	Stethoscope	2	1
	Needle Burner/Destroyer	2	1
	Hospital cot	2	1
	Wheel chair Folding	1	1
	Stretcher on trolley	1	1
	BP Appartus table model	2	1
	Immunization Equipments		
	Hub cutters Manual	5	1
	ILR with stabilizer	1	1
	Deep Freezer with stabilizer	1	1
	Refregirator with stabilizer 165 liter	1	1
	Centrifuge Machine 8 tube	1	1
	Furniture & Others		
	Table for medical officer	2	1
	Chairs medical officer	2	1
	Chairs for Doctors room	5	1
	Table for other staff	5	1
	Chairs for other staff	5	1
	Plastic Chairs	1	1
	Stool	1	1
	Cupboard Full Size	2	1
	Computer Table	2	1


Computer Chair	2	1
Matress coir fome rexin cover 4" (size 3 ft x 6 ft)	2	1
Bed Sheets White	5	1
Bed Sheets Blue	5	1
Bed Sheets Green	5	1
Furniture & Others		
Benches	5	1
Cupboard small	2	1
Linen		
Pillow Cotton	2	1
Pillow cover size 18" x 24" (White, Green, Blue)	4	1
Blanket 60" x 90"	2	1
Blanket woolen baby	2	1
Solapuri Chadar Size 60"x90"	2	1
Towels Navy Blue / Maroon / Brown Honey comb Towl 18"x24"	10	1
Napkins	10	1
Green Cloth Hospital 60"	20	1
Mackintosh	20	1

सदरील साहीत्य सोबत दिलेल्या गुणनिर्देश व गुणवत्तापुर्वक असावे आणि सोबत खालिल दस्ताऐवज सादर करावे.

अं. क्रं.	दस्ताऐवज
१	पॅन कार्ड सत्यप्रत
२	व्हॅट नोंदणी
३	शॉप अॅक्ट लायसन्स
४	बँक डिटेल्स
५	दरपत्रक रेट
६	तिन वर्षातील ताळेबंद पत्रक

सदर दरपत्रक वैद्यकीय अधिकारी आरोग्य केंद्र हर्सुल, औरंगाबाद यांचे नावे सादर करावे. दर सादर करते वेळी प्रती नग दर सादर करावे, सदरील साहित्य आरोग्य केंद्रातील आवश्यकते नुसार मागविण्यात येईल याची संबंधितानी नोद घ्यावी खरेदी बाबतच्या अटी व शर्ती तसेच सर्व अधिकार सदस्य सचिव तथा आरोग्य वैद्यकीय अधिकारी अधिकारी महानगरपालिका औरंगाबाद यांच्या कडे राखून ठेवण्यात आलेले आहेत.

सदर दरपत्रक वै.अ. आरोग्य केंद्र
हर्सुल, येथे पाठवावे.


वैद्यकीय अधिकारी,
नागरी प्राथमिक आरोग्य केंद्र हर्सुल
महानगरपालिका औरंगाबाद

- 4 Technical Specification of IUCD SET

S. No.	Particulars
1	Sims speculum big/med
2	Anterior vag wall retractor 10"
3	Valsellum 10"
4	Ellis forceps 10"
5	Sponge holding forceps 10"
6	Curved artery forceps 6"
7	Straight scissor 4"
8	Uterine sound 10"

Um
MEDICAL OFFICER
CIDCO N-11, Hospital,
Municipal Corporation,
Aurangabad.

Um
Dr. S. S. ...
...
...

-1 Technical Specification of Delivery kit


S.NO.	Particulars	Qty.
1	Delivery Kit Should be CE, ISO 9001:2000, & 13485:2003 certificates	1 1Pc.
2	Trolley, dressing carriage size 76C, long x 46 cm wide and 84 cm high Ref. IS 4769/1968	1 1Pc.
3	Towel, trolley 84 cm x 54 cm	2Pc
4	Gown, operation, cotton	1Pc
5	Cap. Operation, surgeon's 36 x 46 cm	2 Pc
6	Gauze absorbent non-sterile 200 mm x 6 m as per IS: 171/1985.	2Pc
7	Tray instrument with cover 450 mm (L) x 300 mm (W) x 80 mm (H)	1 Pc
8	Macintosh, operation, plastic	2 Pc.
9	Mask, face, surgeon's cap of rear ties: B) Beret type with elastic hem	2 pcs
10	Towel	2 pcs
11	gloves	3 Pairs
12	Cotton wool absorbent non-sterilize 500G	2 Pc
13	Drum, sterilizing cylindrical - 275 mm Dia x 132 mm, ss as per IS: 3831/1979	2Pc
14	Table instrument adjustable type with tray ss	1Pc.


U3
MEDICAL OFFICER
CIDCO H-11, Hospital,
Municipal Corporation,
Aurangabad.

U3m
श्री. डॉ. वि. लक्ष्मण
मुख्य चिकित्सक, सहायक चिकित्सक
सि.टी.टी. अस्पताल,
सि.टी.टी. कॉम्प्लेक्स,
अहमदनगर जिल्हा, महाराष्ट्र

- Technical Specification of GENERAL ITEMS

SNO.	Particulars
1	Pillow Cotton
2	Pillow cover size 18" x 24" (White, Green, Blue)
3	Blanket 60" x 90"
4	Blanket woolen baby
5	Solapuri Chadar Size 60"x90"
6	Patient Dress (Male) Top & Payjama- Top size 36,38,40,42,44,46 & Payjama size 28,30,32,34,36,38,40,42
7	Towels Navy Blue / Maroon / Brown Honey comb Towel 18"x24"
8	Napkins
9	Green Cloth Hospital 60"
10	Mackintosh


MEDICAL OFFICER
CIDCO H-11, Hospital,
Municipal Corporation,
Aurangabad.


डॉ. वी. डी. रावकर
कार्यालय अणुदर रोग वैद्यकीय अधिकारी
सिडको नगर
नगरपालिका अहमदाबाद

EQUIPEMENTS

S.No.	Particular
1	Hammer
2	Tough Depressor
3	Thermometer Dial
4	Thermometer Digital
5	Dressing drum 11"x 9"
6	Baby Receiving Tray Big
7	Steel tray with cover 12"x10"
8	Steel tray with cover 10"x8"
9	Steel dressing jar
10	Steel basin 30cm
11	Steel Stand for basin
12	Steel kidney tray 8"
13	Steel bowl 10"
14	Steel surgical trolley as per specification attached
15	Bed side screen
16	Stethoscope Pediatric
17	Dust bin plastic with cover set of Red Yellow Black
18	Bed Pan S. S.
19	Urine Pot plastic (Male)
20	Urine Pot plastic (Female)
21	Torch

UM
MEDICAL OFFICER
CIDCO H-12, Hospital,
Municipal Corporation,
Aurangabad.

- Technical Specification of Hub Cutter Manual


Hub cutter, Syringe 1) Material use will have The cutting blade of the cutter will be made of stainless (IS 6911 of 1992). The thickness of the blade will not be less than 0.5 mm one side sharpen to cut minimum of 10,000 cuts. All other metal if used in the hub cutter will be made of galvanized carbon steel which will withstand rusting even in rough weather with high humidity coastal region for a minimum of five year (IS 277 of 2003 – galvanized steel sheet specification) The spring if used in the hub cutter will made of steel wire for mechanical spring (oil harden and tempered steel wire) ISI-4454 of 2001 revision part. The sharp container should be made of polypropylene (PP) of not less than 4 mm thick and should be white/translucent. The material should be non-toxic and pyrogenfree , the material will not wear out with normal usage

• Shape : The shape of the container will be such that the lid is tightened with the body with a twist and containing the cut hub along with needle of various size will be $\frac{3}{4}$ full when filled with 500 cut hub with needles. Capacity range : Able to destroy from hub of AD syringe with needles from 18-28 gauge diameter and from 10-25 mm length and fitted with all types of needle fittings, eg. Fixed luer lock luer slip of snap – on Resistance of piecing : Container shall accept 500 cut needle along with the hub when it is $\frac{3}{4}$ th full (filling line). Needle shall not penetrate container walls when filled up to the filling line. The average of forces needed to penetrate samples shall be not less than 15 N.

Tamper-proofing : When contaminated needles are being dropped into containers, it shall remain sufficiently sealed so as to prevent a hand from entering. Needle should not protrude from the container when the box contains 500 needles with hub. Capacity of the container : have sufficient capacity to hold 500, 148-25 gauge, 10-25 mm long needles with hub. Colour : The container shall be translucent white in colour and will allow visual inspection from outer surface to determine when the container is full.

Technique : The device shall be designed to facilitate the location and entry of the needles so that the AD syringe is cut from hub by having: a non-critical angle of the needle relative to the entry point. An entry target area which is recessed, concave and diameter is such that it allows all type of AD syringe available in the market to cut from hub Handling : Able to be carried in one hand and

portable Instructions : Pictorial instructions is major Indian languages to be printed on the surfaces of the container to clearly describe: Use of the hub cutter Biohazard symbol as stipulated in schedule III of the Bio Medical Waste management and Handling rules : Govt supply not for sale


M. V. D. Srinivas
MEDICAL OFFICER
CIBCO Hospital,
Municipal Corporation,
Hyderabad

Technical Specification OF Refrigerator

1	165 Litre
2	Single Dore
3	Energy Consumption 5 Star

Ums
MEDICAL OFFICER
CIDCO H-11, Hospital,
Municipal Corporation,
Aurangabad.